SIEMENS

UROSKOP ACCESS

SP

Replacement of Parts

System

Keyboard Labeling

© Siemens AG 2005
The reproduction, transmission or use of this document or its contents is not permitted without express written authority. Offenders will be liable for damages. All rights, including rights created by patent grant or registration of a utility model or design, are reserved.

Print No.: SPL5-330.841.05.01.02
Replaces: n.a.

English
Doc. Gen. Date: 03.05

Document revision level

The document corresponds to the version/revision level effective at the time of system delivery. Revisions to hardcopy documentation are not automatically distributed.

Please contact your local Siemens office to order current revision levels.

Disclaimer

The installation and service of equipment described herein is to be performed by qualified personnel who are employed by Siemens or one of its affiliates or who are otherwise authorized by Siemens or one of its affiliates to provide such services.

Assemblers and other persons who are not employed by or otherwise directly affiliated with or authorized by Siemens or one of its affiliates are directed to contact one of the local offices of Siemens or one of its affiliates before attempting installation or service procedures.

Table of Contents 3

1 Prerequisites 4

General 4
Prerequisites 4
Special tools/materials 4
Required documents 4
Ordering information 4
Kit contents 4

2 Keyboard Labeling 5

Work steps 5
Final check 6
Customer information 7

General

Prerequisites

NOTE

Actions to be performed are marked with a bullet point.

Text indented below the action to be performed is intended as general information.

Special tools/materials

- Cleaning solution (2/3 water; 1/3 alcohol)

Required documents

- Operating manual

Ordering information

This document is part of the following kit:

- “Keyboard Labeling” (mat. no. 11 56 806)

This kit has to be ordered from CS ML (SAP distribution channel, factory 2050).

Kit contents

Kit 11 56 806 contains the following parts:

Pos.	Quantity	Material no.	Name
1	1	77 57 557	Keyboard labeling strip
2	1	SPL5-330.841.05.01	Replacement of parts; Keyboard labeling

Work steps

- Clean the keyboard above the function keys.

NOTE

Only wipe the keyboard with a clean damp cloth.

Only use a solution of 2 parts water and 1 part of alcohol to moisten the cloth as described in the operating manual.

NOTICE

Keyboard damage!

Do not pour this fluid onto the keyboard because of possible damage.

➪ **Moisten the cloth away from the keyboard.**

- Pull the symbol strip off of the keyboard strip.
- Place the symbol strip above the function keys on the keyboard as marked.

Final check

- Check whether the symbol strip has been placed correctly on the keyboard without air bubbles.

Customer information

n.a.